

2019

Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a Healthcare Work Environment

Brittany Renee Carver

West Virginia University, brcarver@mix.wvu.edu

Follow this and additional works at: <https://researchrepository.wvu.edu/etd>



Part of the [Biology Commons](#), [Occupational Health and Industrial Hygiene Commons](#), [Operations Research, Systems Engineering and Industrial Engineering Commons](#), and the [Other Public Health Commons](#)

Recommended Citation

Carver, Brittany Renee, "Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a Healthcare Work Environment" (2019). *Graduate Theses, Dissertations, and Problem Reports*. 3946.
<https://researchrepository.wvu.edu/etd/3946>

This Thesis is protected by copyright and/or related rights. It has been brought to you by the The Research Repository @ WVU with permission from the rights-holder(s). You are free to use this Thesis in any way that is permitted by the copyright and related rights legislation that applies to your use. For other uses you must obtain permission from the rights-holder(s) directly, unless additional rights are indicated by a Creative Commons license in the record and/ or on the work itself. This Thesis has been accepted for inclusion in WVU Graduate Theses, Dissertations, and Problem Reports collection by an authorized administrator of The Research Repository @ WVU. For more information, please contact researchrepository@mail.wvu.edu.

Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a Healthcare Work Environment

Brittany Carver

Thesis submitted to the Statler College of Engineering and
Mineral Resources at West Virginia University

in partial fulfillment of the requirement for the degree of

Master of Science
in
Industrial Hygiene

Xinijan He, Ph.D., Chair
Steven Guffey, Ph.D.
Anna Allen, MD.

Department of Industrial and Management System
Engineering

Morgantown, West Virginia
Spring 2019

Keywords: respirators, N95, PAPRs, healthcare
Copyright 2019: Brittany Carver

Abstract

Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a healthcare work environment

Brittany Carver

For those working in the healthcare field, personal protective equipment is vital.

Healthcare workers are often first responders when it comes to dealing with patients who are infected or when a disease outbreak occurs. The Occupational Safety and Health Administration (OSHA) requires all healthcare operations to provide employee protection against blood borne pathogens. With the increase of life-threatening disease outbreaks, such as SARS in 2003 and Ebola in 2014, it is of the upmost importance to make sure respirators are best suited for users and acceptable for long term use. It has been shown through previous studies that due to certain factors, respirator user non-compliance is common. Such factors include but are not limited to headaches, thermal discomfort, eye discomfort, interference with duties, etc. By conducting the current study, data was collected on the above factors based off of user opinions of respirators. This study set out to determine if there was significant difference between sixteen different categories, which included rating respirators. The study included four respirators: one filtering face piece respirator (FFR) and three powered air-purifying respirators (PAPR's). Subjects recruited were healthcare workers and had prior knowledge of respirators. Subjects conducted the study at a simulation laboratory and each subject was instructed to wear each of the four respirators and perform a list of tasks common to HCW's. After completing all tasks, subjects completed a questionnaire based on each respirator. Results showed that the N95 mask was preferred over PAPR's in five of sixteen categories. Within the PAPR's themselves, MAXAIR performed slightly better than Airmate, while Versaflo typically was outperformed by all respirators.

Acknowledgements

The author wishes to acknowledge Dr. Xinijan “Kevin” He, Dr. Steven Guffey, Dr. Anna Allen, Huihui Yang, Lauren MacDowell, Dan Summers, and Adam Hoffman for their help with this study.

Table of Contents

Abstract.....	iv
Acknowledgements	v
Table of Contents	vi
List of Figures.....	viii
List of Tables	ix
List of Abbreviations	x
Chapter 1: Introduction.....	1
Chapter 2: Literature Review	4
52.1 Evaluation of Respirators	4
2.2 N95 Efficiency Studies	4
2.3 PAPR Efficiency Studies	6
2.4 Summary.....	7
Chapter 3: Research Design and Methods.....	8
3.1 Objective.....	8
3.2 Selection of Subjects	8
3.3 Experimental Setup.....	8
3.3.1 Instrumentation	9
3.3.2 Tasks to be Performed	10
3.3.3 Parameters for Respirator Evaluation.....	11
3.4 Protocol.....	13
3.5 Data Analysis.....	13
3.5.1 Categorical Data Analysis	13
3.5.2 Questionnaire Data Analysis	14
Chapter 4: Results.....	15
4.1 N95 Evaluation	15
4.2 PAPR Evaluation	15
4.3 PAPR Flow Rate Preferences	18
Chapter 5: Discussion.....	21
5.1 User Evaluation with Categorical Data	21
5.2 PAPR Flow Rate Preferences	21

5.3 Comparing protection, acceptability, and perceived efficiency when conducting low-risk tasks.....	22
5.4 Comparing preference, acceptability, and effectiveness when conducting high-risk tasks.....	22
5.5 Limitations.....	22
5.6 Conclusion.....	23
References	25
Appendix A: Screening Questionnaire Sample.....	27
Appendix B: PAPR Questionnaire Sample.....	28
Appendix C: Raw Data.....	50
C.1 Evaluation Categories	50

List of Figures

Figure 1: N95 Mask.....	1
Figure 2: Types of PAPRs.....	2
Figure 3: 3M™ Versaflo™.....	9
Figure 4: 3M™ MAXAIR CAPR® 710.....	9
Figure 5: 3M™ Air-Mate™	10

List of Tables

Table 1: List of Exercises to be completed when wearing each Respirator.....	10
Table 2: Qualitative evaluation.....	11
Table 3: Evaluation of flow rate and respirator preference.....	12
Table 4: N95 FFR Descriptive results.....	15
Table 5: Versaflo Descriptive results.....	16
Table 6: Max-air Descriptive results.....	16
Table 7: Air-Mate Descriptive results.....	17
Table 8: Kruskal-Wallis Test Results for Categorical Data.....	18
Table 9: Comparing preference, acceptability, and effectiveness when evaluating flow rate.....	18
Table 10: Comparing preference, acceptability, and effectiveness when conducting low-risks tasks.....	20
Table 11: Comparing preference, acceptability, and effectiveness when conducting high-risks tasks.....	20

List of Abbreviations

ANOVA	Analysis of Variance
APF	Assigned Protection Factor
ATD	Aerosol Transmissible Diseases
CA	Clinical Associate
CDC	Center for Disease Control
CFM	Cubic Feet per Meter
CPR	Cardiopulmonary Resuscitation
EKG	Electrocardiogram
EMT	Emergency Medical Technician
EV	Exhalation Valve
FDA	Federal Drug Administration
FFR	Filtering Face Piece Respirators
HCW	Healthcare Worker
HSC	Health Sciences Center
IV	Intravenous
METI	Medical Education Technologies Inc.
MUC	Maximum Use Concentration
NIOSH	National Institute of Occupational Safety and Health
NPPTL	National Personal Protective Technology Laboratory
OSHA	Occupational Safety and Health Administration
PAPR	Powered Air-Purifying Respirator
PEL	Permissible Exposure Limit

PPE	Personal Protective Equipment
RN	Registered Nurse
SAR	Supplied Air Respirator
SI	Speech Intelligibility
SpO2	Blood Oxygen Level
STI	Speech Transmission Index
SWPF	Simulated Workplace Protection Factor
TIL	Total Inward Leakage
WPF	Workplace Protection Factor
WV STEPS	West Virginia Simulation Training and Education for Patient Safety
WVU	West Virginia University

Chapter 1: Introduction

Personal protective equipment, or PPE, is a necessity for those who work in the healthcare field. The typical PPE for healthcare workers includes gloves, goggles or glasses, a face shield, a gown, shoe covers, and a mask or respirator, all of which provide a barrier between the user and any infectious material. While the gloves, goggles, and face shields provide external protection, it is the mask or respirators that will provide protection for a user's respiratory tract (Center for Disease Control, NA). Healthcare work environments pose hazards to workers because they presumably contain hazards such as aerosol transmissible diseases (ATDs).

The Occupational Safety and Health Administration (OSHA) requires all healthcare operations to provide employees protection against blood borne pathogens. In certain settings, engineering and administrative controls do not adequately protect users from airborne droplets, and therefore workers should use a form of respiratory protection. The most common form of respiratory protection is the N95 filtering face piece respirator (FFR), shown in figure 1.



Figure 1: N95 Mask, Left: No valves, Right: N95 with exhalation valve

The most basic N95 respirators are designed to give the user a close facial fit and filtrate airborne particles. Some mask however, are equipped with exhalation valves to reduce heat build-up. The N stands for NOT resistant to oil, and are the only FFR's that have no service life. Other FFR's can be labeled R for resistant to oil, or P for protection against solid or liquid aerosols that may contain oil. The 95 means that this respirator blocks at least 95% of small (0.3 micron) test particles and are considered to exceed the filtration capabilities of the basic facemask. These mask are not compatible with children or people with facial hair. N95 mask are designed for use in industrial and health care settings. In the healthcare field specifically, the N95 respirators will be single-use and are considered a class II device that is regulated by the FDA (US Food and Drug Administration, 2018).

While N95 mask provide protection against some airborne particles, they do not completely eliminate the risk of transmission of diseases. When a higher level of protection is required, then a worker should use a powered air-purifying respirator (PAPR). Defined by OSHA, a PAPR is a respirator that provides protection to the user by filtering out the airborne contaminants and using a battery-powered blower to supply the user with clean air via either a helmet, a tight-fitting respirator, or a loose-fitting hood (Board of Health Sciences Policy, 2015), shown in figure 2.




Figure 2: Types of PAPRs

Over the last decade, there have been several life-threatening disease outbreaks throughout the world. Healthcare workers are often the first responder's when it comes to controlling and maintaining an outbreak, since they are in close proximity with infected patients (Macintyre., Chughtai., Seale., Richards., & Davidson, 2014). In the 2003 outbreak of Severe Acute Respiratory Syndrome (SARS), the Canadian outbreak alone resulted in more than 200 healthcare workers being infected, including three deaths. Likewise, the more recent 2014 Ebola

outbreak resulted in 240 healthcare workers becoming infected and 120 workers died (Moore, D., Yassi, A., & Bryce, E. 2004).). With the increasing rate of disease, it is of utmost importance that healthcare workers wear proper respiratory protection. However, due to comfortability issues, wearing respirators for long periods of time can be difficult.

Chapter 2: Literature Review

2.1: Evaluation of Respirators

There are several factors that determine the degree of protection afforded by a respirator and adequate protection is only achievable when they are selected for the proper tasks, are properly fitted for each user, and adequately maintained (OSHA7). One way to determine the efficiency of an FFR is to look at the assigned protection factor (APF), which is defined as the level of respiratory protection that a class of respirators can achieve when an employer properly follows a respiratory protection program. N95 masks have an assigned APF of 10, while PAPR's have an APF of 25 (OSHA8). Similarly, the maximum use concentration (MUC) can also be used to evaluate the respirator. The MUC is defined as the maximum concentration of a pollutant that an employee will be protected from when using that specific class of respirator. To determine the MUC of a given respirator, the OSHA permissible exposure limit (PEL) of the substance the employee will be exposed to is multiplied by the APF (Steelnack, 2007). Another evaluation factor used when determining the protection of a respirator is the total inward leakage (TIL). It is defined by the amount of contaminated air that leaks through a respirators face seal, valve, gasket, or by filter penetration (Baugh, 2015). Another important test used to determine the efficiency of FFR's is a fit-test, which will assess the type, size, and model of respirator needed for a specific user. The general fit testing procedures given by OSHA include several issues, such as making sure the user properly puts the respirator on, having the user asses comfort based on the position of mask on the nose, room for eye protection, room to talk, and position of mask on the face/cheek. The user must also determine adequacy of the respirator fit based on the strap tension, the fit across the nose bridge, proper size span from nose to chin, and tendency of respirator to move or slip. The user must also perform a seal check by using the positive and negative pressure seal checks given by OSHA, as well as perform a number of test exercises, including deep breathing and body movements (OSHA 29 CFR 1926.10.).

2.2 N95 Efficiency Studies

There have been several studies investigating the effectiveness and protection provided by N95 respirators. One previous study conducted aimed to determine the protection factors of N95 FFR's against particles that are similar in size to bacterial and viral particles, while also investigating the inhalation flow rate. The study measured the concentrations of particles outside and inside an N95 respirator using a sampling system connected to an electrical low pressure

impactor that measures the aerodynamic size of particles ranging from 0.029-10.18 μm . Similarly, particles found in common diseases, such as SARS, H5N1, and *B.anthraxis* range in sizes from 0.08-0.86 μm . Next, the protection factor was calculated by dividing the concentration of particles outside of the mask by the concentration of particles inside the mask. Results showed that the lowest protection factor provided by the N95 FFR occurred when particles were 0.08-0.2 μm . Also, the following shows percentages for the four types of respirators tested, that had a protection factor less than 10: respirator A (13.9%), respirator B (63.9%), respirator C (11.1%), and respirator D (22.2%). Not only have N95 respirators demonstrated a lower efficiency than expected, studies have also investigated other avenues that deem the N95 FFR's unsuitable.

While the tolerability of an N95 FFR is subject to many variables, one of the major problems impacting the effectiveness of respirators is user discomfort. One of the most frequent complaints from N95 FFR users are related to the discomfort from facial and body heat. Thermal discomfort has also been shown to be a main reason for non-compliance when using the N95 FFR. One study focused on the effects of N95 FFR's on core and skin temperatures of users. Subjects were given two N95 FFR's, one with an exhalation valve (EV), and one without. Each subject performed a fit test and then walked on a treadmill for varying amounts of time, the core temperature was measured by an ingestible capsule and skin temperature was monitored with dermal sensors. The results showed minimal increases in core temperature and while skin temperatures of the cheek showed no significant increases at baseline, temperature increased as a function of time. Similarly, cheek temperatures fluctuated between FFR types. (Roberge, 2012).

Likewise, a similar study conducted required twelve subjects to wear three different types of respirators, including an N95 FFR, and two PAPR's. Each subject walked on a treadmill while undergoing physiological response monitoring. Results showed that when conducting low moderate work over a one hour period, there were no significant differences between the one FFR and the four PAPR's tested, in terms of heart rate, respiratory rate, oxygen saturation, transcutaneous CO₂, exertion, facial heat, and body heat. However, respirator dead space heat and humidity as well as facial temperature, were significantly higher for N95 respirators, compared to the other PAPRs (Powell, Kim, & Roberge, 2017). Despite these studies showing that facial and body heat do affect user comfort and acceptability of respirators, other studies have shown contrasting results.

Due to healthcare workers (HCW) needing to be able to communicate clearly with people, especially during public health emergencies, and with hospitals typically being loud environments, speech intelligibility (SI) can become a problem. During the SARS outbreak, 47% of respondents to a health survey from hospital staff reported that wearing respiratory PPE was related to communication interference (Roberge, 2008). Research has shown that interference of FFR's with SI can lead users to be less compliant. Palmiero, Symons, Morgan, and Shaffer performed an assessment of SI while wearing protective facemasks using the Speech Transmission Index (STI). Results showed that when users wore an N95 FFR, their SI interference differed from controls by 17%. The study also suggest that implementing engineering controls and improving designs could be the most effective way to improve SI in

healthcare environments (Palmiero, Symons, Morgan, & Shaffer, 2016). Similarly, another study showed that SI is decreased by at least 1-17% when wearing certain respirators when using the modified rhyme test, and that this is substantial enough to interfere with HCW job tasks (Radonovich Jr., Yanke, Cheng, & Bender, 2009).

Another area that had been documented to interfere with respirator compliance and comfortability is the inducement of headaches after long-term use of FFR's. One study conducted by Lim, et al., set out to determine risk factors associated with the development and impact of headaches among HCW's. There were a total of 212 participants, with 79 of them reporting FFR associated headaches, and 26 of the 79 that reported headaches, also reported a frequency greater than six times a month (Lim, et al., 2006). Another study showed similar results, stating that wearing N95 masks for long durations may induce physiological stress and cause headaches. However, the study suggest this could be due to different temperatures and humidification between the inner and outer areas of the facemasks (Zhu, Lee, Wang, & Lee., 2014).

2.3. PAPR Efficiency Results

While the number of studies investigating the efficiency and comfort factors is in an abundance for N95 respirators, research on PAPR's is less dense. However, there are studies showing that PAPR's provide adequate protection and have little discomfort issues among users.

One study conducted by Koivisto et al, (2015) assessed the program protection factor (PPF), which is the protection a respirator provides in the context of a specific respirator program, provided by a loose-fitting PAPR. The study recruited three subjects and each were required to coat nanoparticles using a liquid flame spray process, while under a ventilation hood. The results showed that the respirators PPF exceeded more than 40×10^3 higher than the original APF of 25. A different study investigated the perceived protection of PAPR's during the SARS outbreak among healthcare workers. Subjects consisted of HCW's who have experience using PAPR's during the SARS outbreak. Results showed that 84% of those surveyed approved the PAPR for protection and preferred it over an N95 FFR when treating SARS patients (Khoo, Leng, Ibrahim, & Lim, 2005). Similarly, another study measured the protection factors of PAPR's, as well as supplied-air respirators (SAR) with hoods and helmets. A total of twelve subjects performed a list of exercises while in a simulated workplace. Results showed that the majority of tested respirators provided a high degree of protection, with the SAR providing less protection than the PAPR's (Cohen, Hecker, Mattheis, Johnson, Biermann, & Foote, 2010).

While PAPR's are considered more comfortable because they supply a cool airflow, reduce breathing effort, are light-weight, easy to use, don't require a fit test, and they cover your head, neck, there are some avenues that allow for discomfort. Some reported disadvantages of PAPR's are the weight of the battery or blower, the battery has to be charged, it can be very noisy, and that it can be awkward for certain tasks (Lenhart, Seitz, Trout, & Bollinger, 2004). One study investigated thermal sensations and comfort of PAPR's. The study recruited twelve subjects and required them to wear five different types of respirators, which included an N95, one tight-fitting full face piece PAPR, two loose-fitting PAPR's, and one elastomeric PAPR.

Each subject was required to walk on a treadmill for a total of one hour while having certain parameters measured. Results showed that the N95 FFR had a higher tympanic forehead temperature and a higher respirator dead space compared to the other PAPR'S. However, the perception of eye dryness was highest among the tight fitting PAPR, and the two loose-fitting PAPR's had a higher rate of facial heat, but this is thought to be due to exercise induced body heat (Powell, Kim, & Roberge, 2017). A different study conducted by Johnson et al., (2007) evaluated sixteen subjects wearing a tight-fitting PAPR while exercising on a treadmill. The PAPR gave different flow rates ranging from 0%-100% of maximum blower capacity of 110 L/min. Results showed that as the flow rate reduced, performance time did as well. Also, differences in breathing comfort and facial thermal comfort were lower when the flow rate was lower, allowing researchers to conclude that higher flow rates allow users to have a better performance time, more facial cooling and respirator comfort (Johnson et al., 2007).

2.4 Summary

While there are a plethora of studies on N95 respirator use, and while there are several studies on PAPR's alone, few studies focus on both types of respirators and furthermore, there is a big literature gap comparing the use of FFR's to PAPR'S. With the amount of protection needed for HCW's when working with patients with serious diseases, and with the amount of increasing disease outbreaks over the years, having the best suitable respirators available to HCW's is vital. Therefore, the purpose of the following study was to evaluate user's perceived acceptability, preference, and effectiveness of four different respirators, and to compare them to each other. Specifically, compare the N95 FFR to the three PAPR's.

Chapter 3: Research Design and Methods

3.1 Objective

This study, in conjunction with others, are part of the efforts of the National Personal Protective Technology Laboratory (NPPTL) at the CDC/NIOSH. The objective of this study was to assess HCW perceptions of comfort, suitability, and protective efficiency of different types of respirators, within a simulated healthcare setting by performing various tasks with varying degrees of risk. Respirators included: the 3M™ particulate respirator 8210 N95 160 EA, MAXAIR CAPR® 710, 3M™ Versaflo™ TR-600-ECK and the 3M™ Air-Mate™. The results of this study were used to compare the N95 FFR to the other three PAPR's in order to determine whether the N95 had similar ratings to the other PAPR's in terms of comfort, suitability, and protective efficiency. Also, we wanted to determine which respirator healthcare workers preferred to use overall.

3.2 Selection of Subjects

There were a total of 16 subjects recruited for this study and participation was voluntary. The purpose of this study, as well as the requirements for participation were emailed to HCW's in selected departments at Ruby Memorial Hospital in Morgantown, WV. Participants included both male (2) and females (14) that worked full-time as one of the following: a physician, registered nurse (RN), emergency medical technician (EMT), respiratory therapists, or clinical associate (CA). Requirements were as follows: between the ages of 18 to 55, non-smokers, no beard (only for N95), English speaking, not pregnant, do not have a systemic disease, and have been trained to use a PAPR. Those wishing to participate contacted Dr. Anna Allen (Clinical Professor, West Virginia University Hospital) where they completed a medical screening questionnaire (Appendix A), and were cleared for respirator use. Based on the medical screening questionnaire, the amount of experience given for each activity listed were totaled and subjects with the highest experience were selected first. Likewise, those with previous PAPR experience were invited to participate before those who had no experience.

3.3 Experimental Setup

In order to replicate a healthcare work environment as much as possible, subjects carried out the lists of tasks at the West Virginia Simulation Training and Education for Patient Safety (WVSTEPS) center at the campus of West Virginia University located in Morgantown, WV. Each simulation room consists of a METI human patient simulator, which is a life-like mannequin capable of producing breathes, blinking eyes, a pulse, a heartbeat, an airway that can mimic different levels of pulmonary obstruction, expires CO₂, and can generate muscle twitches. The types of physiological situations needed for each unique learning situation or study can be manipulated in the computer room via the program called Laerdal Learning Application. In addition to the mannequin, the rooms are also equipped with monitoring devices, suction, stethoscopes, intravenous (IV) bags, needles, syringes, bed pans, EKG monitor, and an extra prosthetic arm equipped with fake blood. The same room and mannequin were used for all 16

subjects, in which the mannequin is a male, 85 pounds, and 5.9 feet tall. The temperature was set at a constant 70° Fahrenheit.

3.3.1 Instrumentation

A total of four respirators were used, one type of N95 FFR and three different types of PAPR's. All respirators used in this study are known to be used in healthcare settings.

- 1) 3M™ particulate respirator 8210 N95 160 EA, shown in figure 1, the system is a one-time use disposable mask that has a two-strap design with welded dual point attachment, a cushioned nose area, is lightweight, and has an advanced electrostatic media.
- 2) 3M™ Versaflo™ TR-600-ECK (loose-fitting), shown in figure 3, consists of a motor (TR-602N), a battery (TR-971), easy clean belt (TR-627), airflow indicator (TR-971), a length adjusting breathing tube (BT-30), a chemical cartridge, and a filter. The standard airflow is 6.7 cfm and the system weights 4.5 lbs.



Figure 3: 3M™ Versaflo™ TR-600-ECK

- 3) MAXAIR CAPR® 710 System DLC-double shroud (loose-fitting), shown in figure 4, consists of a helmet (2065-03), battery (2000-36), charger (2600-01), belt (2000-76), and filter (2160-10). The airflow is set a 6 cfm and the total weight is 2.5 lbs.



Figure 4: MAXAIR CAPR® 710

- 4) 3M™ Air-Mate™ Belt-Mounted High Efficiency PAPR, shown in figure 5, consists of a hood assembly (520-03-63R01), battery (007000015R01), high efficiency filter (451-02-01R0), nylon waist belt (021-14-00R01), and an airflow indicator (021-14-00R01). The airflow can be set at different levels and the total weight is 3 lbs.



Figure 5: Air-Mate™

3.3.2. Tasks to Be Performed

In order for the subjects to be able to completely evaluate each respirator, a set of tasks that are common in the healthcare field were to be performed by each subject with all four respirators. After collaborating with the simulation lab personnel, a list comprised of twelve tasks was developed and are shown in figure 6, along with a description of each. It was estimated that the total time for each subject to complete all tasks and the questionnaire would take forty-five minutes to one hour. Although, most subjects completed all tasks within fifteen minutes, and along with the questionnaire the total time for each subject was around thirty minutes.

Table 1: List of Exercises to Be Completed When Wearing Each Respirator

Tasks	Description
Setting up bedding	Fold/Position the sheets and blankets, make sure patient has a pillow, organize room
Positioning Patient	Adjust patient by lowering or raising the bed
Asses Patient	Use a stethoscope to asses patients breathing/heartrate, take pulse, look at eyes, nose, throat, and ears
Hook up monitoring devices	Connect the electrodes to the patient
Insert IV	Identify vein, use Chloroprep to clean the area, puncture vein using needle/ IV catheter
Draw blood	Hook syringe into catheter, pull back on syringe to collect blood

Start IV equipment	Attach an IV bag to the IV catheter via tubing and start the equipment
Insert and set up Foley	Clean area, lubricate tubing, insert tubing into area, inflate the tubing via water-filled syringe
Take urine sample	Using the urine collection bag, collect a small sample of urine
Turn and bathe patient	Remove patient's clothing, wash patient using wash basing and washcloth. When finished, redress patient with new gown
Suctioning	Turn on the suctioning machine, insert the suction tube into the patients mouth, go to all quadrants of the mouth
Perform CPR	Perform at least two rounds of CPR

3.3.3. Parameters for Respirator Evaluation

After completion of all twelve tasks, each subject completed a questionnaire, which is shown in Appendix B. Qualitative evaluation of respirators was determined by choosing which number best expressed how the user felt about that specific parameter, these are shown in table 2. Evaluation of flow rate and respirator preference when performing low-risk and high-risk tasks are shown in table 3.

Table 2: Qualitative Evaluation

Category	Scale
General Comfort	1- Very poor 2- Poor 3- Acceptable 4- Good 5- Very good 6- Excellent
Inspiratory/Expiratory effort	1- Not noticeable 7 – Noticeable
Overall breathing discomfort	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Facial and body heat	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Overall thermal comfort	1 – Coldest you've ever been 5 – Neither cold nor hot 10 – Hottest you've ever been
Pressure or pain	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Eye discomfort	1- Not at all 2- Very slightly

	3- Slightly 4- Somewhat high 5- High 6- Unbearable
Interference with wearing glass/goggles/contact lenses	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Clear line of vision	1- Very poor 2- Poor 3- Acceptable 4- Good 5- Very good 6- Excellent
Difficulty putting on	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Difficulty to operate	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Mechanical interference with duties	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable
Exertion	6 – No exertion at all 20 – Maximal exertion
Hours you could wear this respirator continuously?	Written response
Perceived efficiency against biological hazards	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat High 5- High 6- Complete
Overall Assessment	1- Not at all 2- Very slightly 3- Slightly 4- Somewhat high 5- High 6- Unbearable

Table 3: Evaluation of Flow Rate and Respirator Preference

Question	Type of response
When performing the following tasks (deemed low risks), do you prefer a lower flow rate for the following PAPR?	Yes or No for each respirator

When performing the following tasks (deemed low risks), do you think a lower flow rate is more acceptable?	Yes or No for each respirator
When performing the following tasks (deemed low risks), do you think a lower flow rate is more effective?	Yes or No for each respirator
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you prefer to use?	Yes or No for each respirator
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more acceptable?	Yes or No for each respirator
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more effective?	Yes or No for each respirator
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you prefer to use?	Yes or No for each respirator
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more acceptable?	Yes or No for each respirator
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more effective?	Yes or No for each respirator
Nausea	Yes or No for each respirator
Dizziness of difficult concentrating	Yes or No for each respirator
Unusual smell or odor	Yes or No for each respirator
Eye irritation	Yes or No for each respirator

3.4. Protocol

Once subjects arrived at the WV STEPS center, they were directed to the donning room where they would complete a signed consent. The risk associated with the experiment, as well as an overview of what was expected was given to the subject. Next, the subject was taken down the hall to the simulation room and a brief orientation of the room and materials were given to the subject to ensure familiarity when performing the list of tasks.

Upon returning to the donning room, subjects would then put on the first respirator assigned. In order to reduce bias, the order in which the subjects wore the types of respirators were randomized using Microsoft Excel. Subjects were taken to the simulation room with their respirator on and running. Once beginning the tasks, a timer was set to twenty minutes. There was always at least one investigator in the room with the subject to oversee, sometimes there were extra investigators present.

Once all tasks were completed and the room was put back in order, subjects returned to the donning room to remove all equipment. As subjects answered the questionnaire, the investigator would clean all materials before returning them to their proper area. The subject would then continue through the same steps for the next three respirators. Upon completion of the study, subjects were compensated and given receipts for tax purposes.

3.5 Data Analysis

All questionnaire data was transferred to Microsoft Excel and separated into two categories: categorical and questionnaire. The categorical data was analyzed using SAS JMP software, while the questionnaire data was analyzed using Microsoft Excel.

3.5.1. Categorical Data Analysis

There were a total of sixteen categories, each having four groups, one for each respirator. Due to the data being ordinal, a normal distribution couldn't be assumed and non-parametric test

were performed using SAS JMP software. To compare the three different PAPR's to the N95 FFR, non-parametric test, specifically a Kruskal-Wallis test was performed. An alpha level was set at 0.05, as well as 0.1 for comparison efforts.

3.5.2. Questionnaire Data Analysis

There were a total of thirteen questions, all of which consisted of only yes or no responses. The questions compared preference, acceptability, and effectiveness for flow rates, high-risk tasks, and low-risk tasks. Also included were symptoms such as nausea, dizziness, unusual smell or odor, and eye irritation. Using Microsoft excel, percentages were calculated for all answers.

Chapter 4: Results

4.1 N95 Evaluation

Each of the sixteen subjects evaluated the N95 FFR for sixteen categories and the summary is shown in Table 4. The mean, standard deviation, minimum and maximum values were calculated for each category.

Table 4: N95 FFR Descriptive Results

Category	Mean	Standard Deviation	Minimum	Maximum
General Comfort	2.58	1.177	1	5
Inspiratory/Expiratory effort	3.52	1.589	1	7
Overall breathing discomfort	3.16	1.293	1	5
Facial and body heat	3.84	1.293	1	6
Overall thermal comfort	6.55	1.410	3	10
Pressure or pain	2.81	1.223	1	5
Eye discomfort	1.71	1.160	1	4
Interference with wearing glasses/goggles/contacts	2.52	1.951	1	8
Clear line of visions	4.65	1.226	2	6
Difficulty putting on	1.74	1.064	1	5
Difficulty to operate	1.16	0.454	1	3
Mechanical interference with duties	1.39	0.803	1	4
Exertion	9.65	2.961	6	16
How many hours could you wear this respirator continuously?	1.79	2.232	0	12
Perceived efficiency against biological hazards	3.52	1.338	1	6
Overall assessment	3.87	1.176	1	6

Please note that all categories were scaled from one to six, with one being the best possible choice. There are a few exceptions, general comfort and clear line of vision were scaled from one to six, with one being the worst possible choice. Overall thermal comfort was scaled from one to ten with one being the coldest and ten being the hottest. Exertion was scaled from six to twenty, with six being the best choice.

4.2 PAPR Evaluation

Each of the sixteen subjects evaluated all three PAPR'S for each of the sixteen categories and the summary for each PAPR are shown in Tables 5-7. The mean, standard deviation, minimum and maximum values were calculated for each category. The same scales were identical to those used for the N95 FFR.

Table 5: Versaflo Descriptive Results

Category	Mean	Standard Deviation	Minimum	Maximum
General Comfort	2.94	1.209	1	5
Inspiratory/Expiratory effort	2.32	1.351	1	5
Overall breathing discomfort	2.06	1.237	1	5
Facial and body heat	2.65	1.496	1	6
Overall thermal comfort	5.45	1.895	1	10
Pressure or pain	2.00	1.317	1	5
Eye discomfort	1.39	0.667	1	3
Interference with wearing glasses/goggles/contacts	1.96	1.186	1	5
Clear line of visions	3.48	0.926	2	6
Difficulty putting on	3.23	1.055	1	5
Difficulty to operate	2.55	1.434	1	5
Mechanical interference with duties	3.29	1.321	1	6
Exertion	9.48	3.395	6	17
How many hours could you wear this respirator continuously?	1.99	1.567	0.2	6
Perceived efficiency against biological hazards	3.94	1.263	2	6
Overall assessment	3.48	1.546	1	6

Table 6: Max-air Descriptive Results

Category	Mean	Standard Deviation	Minimum	Maximum
General Comfort	3.74	1.094	2	6
Inspiratory/Expiratory effort	1.58	0.848	1	5
Overall breathing discomfort	1.29	0.529	1	3
Facial and body heat	2.19	0.980	1	5
Overall thermal comfort	5.39	1.256	1	8
Pressure or pain	1.94	1.153	1	5
Eye discomfort	1.45	0.995	1	5
Interference with wearing glasses/goggles/contacts	2.00	1.382	1	5
Clear line of visions	4.03	1.354	1	6
Difficulty putting on	2.35	0.985	1	5
Difficulty to operate	1.65	0.755	1	3
Mechanical interference with duties	2.68	1.045	1	5
Exertion	8.13	1.910	6	13

How many hours could you wear this respirator continuously?	2.47	1.751	0.15	8
Perceived efficiency against biological hazards	3.42	1.409	1	6
Overall assessment	2.84	1.068	1	5

Table 7: Air-Mate Descriptive Results

Category	Mean	Standard Deviation	Minimum	Maximum
General Comfort	3.77	1.055	1	6
Inspiratory/Expiratory effort	1.90	0.908	1	4
Overall breathing discomfort	1.61	0.844	1	4
Facial and body heat	2.00	1.000	1	4
Overall thermal comfort	4.87	0.991	2	7
Pressure or pain	1.58	0.720	1	3
Eye discomfort	1.29	0.783	1	4
Interference with wearing glasses/goggles/contacts	1.87	1.140	1	4
Clear line of visions	4.03	1.140	2	6
Difficulty putting on	2.77	0.990	1	5
Difficulty to operate	2.13	0.885	1	4
Mechanical interference with duties	2.52	1.151	1	5
Exertion	8.39	2.552	6	17
How many hours could you wear this respirator continuously?	2.08	1.106	0.2	4
Perceived efficiency against biological hazards	3.68	1.249	1	6
Overall assessment	2.97	1.140	1	6

All data was reformatted for further analysis using SAS JMP statistical software and non-parametric test were conducted. The α -level for the Kruskal-Wallis test were set at 0.1 and 0.5, and all categories from the questionnaire were included. The results from the test are shown below in table 8.

Table 8: Kruskal-Wallis Test Results for Categorical Data

Category	Chi-squared Result	p-value	Hypothesis results ($\alpha<0.05$)	Hypothesis results ($\alpha<0.1$)
General Comfort	11.39092	0.0442	Reject null (p<0.05)	Reject null (p<0.1)
Inspiratory/expiratory effort	0.071705	0.9994	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Overall breathing discomfort	0.261475	0.9922	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Facial and body heat	4.312439	0.3654	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Overall thermal comfort	10.24852	0.1146	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Pressure or pain	8.303087	0.0401	Reject null (p<0.05)	Reject null (p<0.1)
Eye discomfort	0.573388	0.9025	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Interference with wearing glasses/goggles/contacts	25.781	0.0011	Reject null (p<0.05)	Reject null (p<0.1)
Clear line of vision	8.862527	0.0646	Fail to reject null (p>0.05)	Reject null (p<0.1)
Difficulty putting on	3.16049	0.5313	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Difficult to operate	1.739704	0.7835	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Mechanical interference with duties	3.336675	0.5031	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
Exertion	8.591805	0.2833	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)
How many hours could you wear this respirator continuously?	28.2711	0.0009	Reject null (p<0.05)	Reject null (p<0.1)
Perceived efficiency against biological hazards	9.641867	0.0860	Fail to reject null (p>0.05)	Reject null (p<0.1)
Overall assessment	2.326425	0.6760	Fail to reject null (p>0.05)	Fail to reject null (p>0.1)

When $\alpha<0.05$, the Kruskal-Wallis test showed that a total of four categories showed significant results. Those categories include: general comfort, pressure or pain, interference with wearing glasses/goggles/contacts, and how many hours could you wear this respirator continuously. When $\alpha<0.1$, six categories showed significant results including the same four as when $\alpha<0.05$, along with the categories for clear line of vision and perceived efficiency against biological hazards.

4.3 PAPR Flow Rate Preference

This set of data evaluated the preference, acceptability, and effectiveness of flow rates, as well as respirator preference when conducting high-risk and low-risk tasks. Also, symptoms of nausea, dizziness, odor, and eye irritation were also evaluated. The results are shown as frequency percentages in tables 9-11.

Table 9: Comparing preference, acceptability, and effectiveness of flow rates

Question	Respirator	Yes	No
When performing the following tasks (deemed low risks), do you prefer a lower	N95	NA	NA
	Versaflo	50%	50%
	MAXAIR	38%	63%
	AIRMATE	50%	50%

flow rate for the following?			
When performing the following tasks (deemed low risks), do you think a lower flow rate is more acceptable?	N95	NA	NA
	Versaflo	50%	50%
	MAXAIR	50%	50%
	AIRMATE	63%	38%
When performing the following tasks (deemed low risks), do you think a lower flow rate is more effective?	N95	NA	NA
	Versaflo	38%	63%
	MAXAIR	38%	63%
	AIRMATE	50%	50%

The N95 couldn't be evaluated in terms of flow rate due to being an FFR not a PAPR. While results for Versaflo and MAXAIR were each 50% for both yes and no answers, in terms of preferring the PAPR have a lower flow rate, most of the subjects (63%) preferred that the MAXAIR did not have a lower flow rate. Similarly, the Versaflo and MAXAIR were each 50% for both yes and no answers in terms of whether a lower flow rate is more acceptable. However, the majority (63%) of subjects thought it was acceptable for a lower flow rate for the AIRMATE PAPR. In addition, the AIRMATE has a 50% yes or no answer rate when determining whether a lower flow rate is more effective. However, 63% of subjects for both the Versaflo and the MAXAIR thought a lower flow rate would not be more effective.

Table 10: Comparing preference, acceptability, and effectiveness when conducting low-risks tasks

Question	Respirator	Percentage
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you prefer to use?	N95	25%
	Versaflo	0%
	MAXAIR	38%
	AIRMATE	25%
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more acceptable?	N95	50%
	Versaflo	0%
	MAXAIR	25%
	AIRMATE	25%
	N95	25%

When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more effective?	Versaflo	0%
	MAXAIR	63%
	AIRMATE	13%

When determining which respirator the subjects preferred to use when conducting low-risk tasks, most subjects chose MAXAIR (38%), while 0% chose AIRMATE. When determining which respirator was more acceptable, subjects preferred the N95 FFR (50%), followed by both MAXAIR and AIRMATE at 25%. In addition, when determining which respirator was most effective, subjects preferred MAXAIR 63% more than other respirators, while no subjects (0%) chose Versaflo.

Table 11: Comparing preference, acceptability, and effectiveness when conducting high-risks tasks

Question	Respirator	Percentage
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you prefer to use?	N95	13%
	Versaflo	0%
	MAXAIR	50%
	AIRMATE	38%
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more acceptable?	N95	13%
	Versa flow	13%
	MAXAIR	50%
	AIRMATE	25%
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more effective?	N95	13%
	Versa flow	13%
	MAXAIR	63%
	AIRMATE	13%

Similar to the low-risk tasks data, when determining which respirator the subjects preferred to use when conducting high-risk tasks, most subjects chose MAXAIR (50%), while 0% chose Versaflo. In terms of acceptability, most subjects preferred MAXAIR (50%), followed by AIRMATE (25%). Additionally, when evaluating the most effective respirator, subjects preferred MAXAIR (63%), while the other three respirators were all 13%.

Chapter 5: Discussion

5.1 User evaluation with Categorical Data

Data from the current study indicate that, when comparing the means of the N95 FFR mask to each of the three PAPR's individually, all three PAPR's were rated higher in eleven of the sixteen categories. The five categories in which the N95 was more preferred over the PAPR's were as follows: clear line of vision, difficulty to put on, difficulty to operate, mechanical interference with duties, and perceived efficiency against biological hazards. This is likely due to the fact of the simple design of the N95 when compared to an entire PAPR setup with the hood, belt, battery, etc. It was anticipated that the N95 would rank less than all PAPR's in terms of protection against biological hazards due to the fact that the PAPR equipment covers more of the body and supplies air, however it was not rated lowest in this category. It is worth mentioning that the only PAPR of the three that was rated lower than the N95 in terms of perceived efficiency against biological hazards was MAXAIR, the other two PAPR's were preferred over the N95 in this category. It is also worth mentioning the results of the PAPR's themselves within each category. For the following six categories: inspiratory/expiratory, overall breathing, exertion, how many hours could you wear this respirator continuously, and overall assessment, MAXAIR was rated best followed by Airmate and then Versaflo. When evaluating the following categories: general comfort, facial and body heat, overall thermal comfort, and pressure or pain, Airmate was rated best followed by MAXAIR and then Versaflo. The last three remaining categories all had a mixture of results. For eye discomfort, Airmate was most preferred to not interfere, followed by Versaflo and then MAXAIR. When looking at whether the PAPR's interfered with wearing goggles/glasses/contacts, Airmate had no interference, followed by Versaflo, and then MAXAIR showing the most interference. Lastly, in terms of perceived efficiency of protection against biological hazards, Versaflo was thought to be more protective, followed by Airmate and MAXAIR.

When evaluating the results of the Kruskal-Wallis test, when $\alpha < 0.05$, only four of sixteen categories showed significant results when comparing the results of all three PAPRS combined to the N95. These four categories were general comfort, pressure or pain, interference with wearing glasses/goggles/contacts, and how many hours could you wear this respirator continuously. Also, when $\alpha < 0.1$, 6 of the sixteen categories were significant, the same four as previously mentioned, along with clear line of vision and perceived efficiency against biological hazards. This is most likely due to the variation in the different types of PAPR's, since we combined the data from all three PAPR's. For instance, MAXAIR and AIRMATE were each preferred in almost all categories over the Versaflo.

5.2 PAPR Flow Rate Preference

When asked whether subjects preferred a lower flow rate for any of the PAPR's, results were mixed, with 50% answering yes and 50% answering no. The only results differing from the median was MAXAIR, in which 63% answered no that they did not prefer a lower flow rate. When asked if subjects thought a lower flow rate is more acceptable for any of the PAPR's,

results were similar to the first question, in which 50% for both yes and no. However, Airmate showed 63% did actually agree that a lower flow rate would be more acceptable. Also, when asked if a lower flow rate would be more effective, while Airmate showed the trending 50% results, both the MAXAIR and Versa flow showed that 38% of subjects did think a lower flow rate would be more effective. With most results showing that the majority of subjects were mixed about how they felt towards having a lower flow rate, this could be due to the subjects not having a chance to really adjust or choose which flow rate they used prior to conducting the tasks. Also, subjects could be confused about the question or wording of the question, as well as not have been thinking about the flow rate while conducting tasks.

5.3 Comparing preference, acceptability, and effectiveness when conducting low-risk tasks

When asked which type of respirator out of the four used that the subject would prefer, most subjects (38%) preferred MAXAIR, while none preferred Versaflo. This could be due to MAXAIR being lightweight (2.5 lbs.), and loose-fitting, Also, a lot of subjects expressed that the helmet fit more comfortably than others because it was more cushioning and the straps distributed the weight of the helmet better. Also, the Versaflo could have been the least preferred to use due to it being the heaviest at 4.5 lbs. When asked which respirator subjects thought was more acceptable, half (50%) answered that the N95 was most acceptable. This could be due the lightweight, easy to use, and easy to put on aspects of the mask. Again, Versaflo rated the least with 0% thinking it is acceptable over the other respirators. When subjects answered which respirator they thought was most effective, more than half (63%) answered that they thought MAXAIR was most effective. This was followed by the N95, then Airmate, and lastly with Versaflo. This could be due to the fact that the MAXAIR is very comfortable, has a fully covering helmet, and is lightweight.

5.4 Comparing preference, acceptability, and effectiveness when conducting high-risk tasks

When asked which respirator the subjects preferred to use when conducting high-risk tasks, 50% preferred MAXAIR, followed by 38% suggesting Airmate, with 13% answering N95, and with 0% preferring Versaflo. These results could be due to the fact that The MAXAIR and Airmate are the lighter weight PAPR's and are more comfortable then Versaflo. Also, users could prefer to use the N95 over the Versaflo because it is easy to use, easy to put on, and lighter. When asked which respirator was most acceptable, MAXAIR was most acceptable at 50%, followed by Airmate at 25%. Both the N95 and Versaflo showed 13% of subjects thinking it was acceptable. Similar to previous results, subjects tend to like the lighter weight designs of MAXAIR and Airmate. When asked which respirator was most effective, as with the other trends, MAXAIR was rated highest at 63%, however, all other three respirators were equally rated at 13% each. This could be due to subjects thinking more about the hazards of high-risk tasks and wanting more protection that is durable but yet comfortable.

5.5. Limitations

There were a few limitations to this study, with the most simplistic being that the sample size was rather small. Increasing the sample size to perhaps 50 subjects would allow for more

variability and could show more significant results. Also, within the small sample size, there were only two groups of healthcare workers, it would be nice to have some physician's to add to the diversity of the group and for comparisons reasons. Typically, registered nurses have the most experience with PAPR's due to them having the most direct contact with patients so adding more physicians could change the results.

Another limitation could be that due to some of the subjects having been working in the field for a long period of time, they were so accustomed to their duties that they finished all tasks very quickly compared to some other subjects. By adding more time that each subject would wear each respirator, either by adding more or different types of tasks, or by collecting data on multiple occasions, would allow for more accurate feel of what the respirator would be like during the real healthcare setting.

Another possible limitation could have been that subjects could have been confused or not careful when answering the ratings for each category on the questionnaire because the rating scale changed several times as you go from category to category. However, all subjects were told to ask questions if they did not understand a question.

5.6. Conclusion

In the study conducted by Lee, Grinshpun, & Reponen (2008), their results showed that an N95 mask is not considered adequate protection for particles found in common diseases. Also, OSHA states that while an N95 mask does protect against airborne particles, when a higher degree of protection is required, a PAPR is best. However, in comparison to these previous studies, the current study showed that while the Airmate and Versaflo were perceived as having a higher protection against biological hazards than the N95, the MAXAIR PAPR was actually thought to be slightly less protective than the N95.

With facial and body heat discomfort being reported as one of the main reasons for non-compliance with respirators, it's important to investigate the ratings of this category. Studies conducted by Roberge (2012) showed that the temperature of the facial skin while wearing an N95 masks increased significantly over time. Likewise, another study conducted showed that in comparison to respirators, N95 masks are shown to have higher facial temperatures (Powell, Kim, & Roberge, 2017). Results of this study showed that, similar to other studies, PAPR's were rated to have less thermal discomfort in comparison to the N95 FFR.

Another commonly reported issue with respirators was SI, with the total SI decreasing up to 17%. When asking subjects for the current study about their interference with their duties however, the N95 had less interference with duties than all PAPR's. This could be due to the fact that the question the subjects answered on the questionnaire asked if there was any mechanical interference with duties, as opposed to specifically asking if SI was effected. Also, the sounds from the batteries on the PAPR's could have played a role in the reduction on SI. Similar to some of the reported disadvantages of PAPR's, such as a heavy battery, noise from blower, and having to charge a battery, several subjects in the current study expressed similar views (Lenhart, Seitz, Trout, & Bollinger, 2004).

Overall, out of the four respirators used, subjects only preferred the N95 FFR over the PAPR's in five of sixteen categories. With these categories being clear line of vision, difficulty putting on, difficulty operating, and mechanical interference with duties are all sensible due to the fact that the N95 is the smallest, lightest, and easiest to use. However, when it came to the other eleven tasks, PAPR's were more preferred. More specifically, MAXAIR was typically the most preferred PAPR, followed by Airmate and then Versaflo. These results could be due to subjects being familiar with MAXAIR, it being the lighter weight PAPR, also the helmet design was liked among subjects. The results of the current study can be used to help fill the literature gap on respirator use and help manufacturers design respirators to be better fit for HCW's or users. Further research should aim to address some of the following aspects of respirator use: whether respirators play a role on physiological aspects such as heartbeat, body temperature, breathing, etc. Also, further studies should use other means of measurement than solely a questionnaire, such as using a thermometer to measure temperature or implementing other devices to measure heartbeat, blood pressure, number of breaths, etc. These types of studies can only further the efforts of providing the best and most adequate protection for healthcare workers, especially in the face of a disease outbreak.

References

1. Baugh, L. (2015). Total Inward Leakage Testing for Respirators. *International Safety Equipment Association*. Retrieved from <https://safetysafetyequipment.org/knowledge-center-items/total-inward-leakage-testing-for-respirators/>
2. Board on Health Sciences Policy; Institute of Medicine. (2015). The Use and Effectiveness of Powered Air Purifying Respirators in Health Care: Workshop Summary. *National Academic Press*. Retrieved from <https://www.ncbi.nlm.nih.gov/books/NBK294223/>
3. Center for disease control. (NA). Guidance for selection and use of personal protective equipment (PPE) in healthcare settings. [Powerpoint slides]. Retrieved from <https://www.cdc.gov/HAI/pdfs/ppe/PPEslides6-29-04.pdf>
4. Cohen, H.J., Hecker, L.H., Mattheis, D.K., Johnson, J.S., Biermann, A.H., & Foote, K.L. (2010). Simulated Workplace Protection Factor Study of Powered Air-Purifying and Supplied Air Respirators. *American Industrial Hygiene Association*, 62(5), 595-604.
5. Johnson, A.T., Mackey, K.R., Scott, W.H., Koh, F.C., Chiou, K.Y.H., & Phelps, S.J. (2007). Exercise Performance While Wearing a Tight-Fitting Powered Air Purifying Respirator with Limited Flow. *Journal of Occupational and Environmental Hygiene*, 2(7), 368-373.
6. Khoo, K.L., Leng, P.H., Ibrahim, I.B., * Lim, T.K. (2005). The changing face of healthcare worker perceptions on powered air-purifying respirators during the SARS outbreak. *Respirology*, 10(1), 107-110.
7. Koivisto, A.J., Aromaa, M., Koponen, I.K., Fransman, W., Jensen, K.A., Makela, J.M., & Hameri, K.J. (2015). Workplace performance of a loose-fitting powered air purifying respirator during nanoparticle synthesis. *Journal of Nanoparticle Research*, 17, 177.
8. Lee, S., Grinshpun, S.A., & Reponen, T. (2008). Respiratory Performance Offered by N95 Respirators and Surgical Masks: Human Subject Evaluation with NaCl Aerosol Representing Bacterial and Viral Particle Size Range, *The Annals of Occupational Hygiene*, 52(3), 177-185. <https://doi.org/10.1093/annhyg/men005>
9. Lenhart, S.W., Seitz, T., Trout, D., & Bollinger, N. (2004). Issues Affecting Respirator Selection for Workers Exposed to Infectious Aerosols: Emphasis on Healthcare Settings. *Applied Biosafety*, 9(1), 20-36.
10. Lim, E.C.H., Seet, R.C.S., Lee, K.H., Wilder-Smith, E.P.V., Chuah, B.Y.S., Ong, B.K.C. (2006). Headaches and the N95 face-mask amongst healthcare providers. *Acta Neurologica Scandinavica*, 113(3), 199-202.
11. Macintyre, C.R., Chughtai, A.A., Seale, H., Richards, G.A., & Davidson, P.M. (2014). Respiratory protection for healthcare workers treating Ebola virus disease (EVD): Are facemasks sufficient to meet occupational health and safety obligations?. *International Journal of Nursing Studies*, 51(11), 1421-1426. Retrieved from <https://www.sciencedirect.com/science/article/pii/S002074891400234X>
12. Moore, D., Yassi, A., & Bryce, E. (2004). Protecting the Faces of Health Care Workers: Knowledge Gaps and Research Priorities for Effective Protection Against Occupationally-Acquired Respiratory Infectious Diseases. *The Change Foundation*. Retrieved from

<http://www.phsa.ca/Documents/Occupational-Health-Safety/ReportProtectingtheFacesofHealthcareWorkers.pdf>

13. Palmiero, A. J., Symons, D., Morgan, J. W., & Shaffer, R. E. (2016). Speech intelligibility assessment of protective facemasks and air-purifying respirators. *Journal of occupational and environmental hygiene*, 13(12), 960-968.
14. Powell, J. B., Kim, J. H., & Roberge, R. J. (2017). Powered air-purifying respirator use in healthcare: Effects on thermal sensations and comfort. *Journal of occupational and environmental hygiene*, 14(12), 947-954.
15. Radonovich Jr., L.J., Yanke, R., Cheng, J., & Bender, B. (2009). Diminished Speech Intelligibility Associated with Certain Types of Respirators Worn by Healthcare Workers. *Journal of Occupational and Environmental Hygiene*, 7(1), 63-70.
16. Roberge, R.J. (2008). Effect of surgical masks worn concurrently over N95 filtering facepiece respirators: extended service life versus increases user burden. *Journal of Public Health Management*, 14(2), 19-26. doi: 10.1097/01.PHH.0000311904.41691.fd.
17. Roberge, R., Benson, S., & Kim, J. (2012). Thermal Burden of N95 Filtering Facepiece Respirators, *The Annals of Occupational Hygiene*, 56(7), 808-814.
<https://doi.org/10.1093/annhyg/mes001>
18. US Food & Drug Administration (FDA). (2018). Masks and N95 Respirators. Retrieved from <https://www.fda.gov/medicaldevices/productsandmedicalprocedures/generalhospitaldevicesandsupplies/personalprotectiveequipment/ucm055977.htm>
19. Occupational Safety and Health Administration. **NIOSH Guide to the Selection and Use of Particulate Respirators Certified Under 42 CFR 84** [DHHS (NIOSH) Publication No. 96-101]. Retrieved from https://www.osha.gov/dts/osta/otm/otm_viii/otm_viii_2.html
20. Occupational Safety and Health Administration. 63 FR 1152, Jan. 8, 1998; 63 FR 20098, April 23, 1998; 71 FR 16672, April 3, 2006; 71 FR 50187, August 24, 2006; 73 FR 75584, Dec. 12, 2008; 76 FR 33606, June 8, 2011]. Retrieved from <https://www.osha.gov/laws-regs/regulations/standardnumber/1910/1910.134>
21. Occupational Safety and Health Administration. 29 CFR 1926.10.
22. Steelnack, J.(2007). Assigned Protection Factor (APF) Table Added to OSHA’s Respiratory Protection <https://www.aiha.org/aihce07/handouts/rt243steelnack.pdf> Standard 29 CFR 1910.134. [Powerpoint presentation].
23. Zhu, J.H., Lee, S.J., Wang, D.Y., & Lee, H. (2014). Effects of long-duration wearing of N95 respirator and surgical facemask: a pilot study. *Journal of Lung, Pulmonary, & Respiratory Research*, 1(4), 97-100.

Appendix A: Screening Questionnaire Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a healthcare work environment

BASIC INFORMATION

PLEASE PRINT AND COMPLETE ALL ENTRIES			
NAME (LAST -- FIRST -- MIDDLE INITIAL)		JOB TITLE	
SEX (M or F)	Phone Number	Email Address	YEARS OF WORK EXPERIENCE
Have you used a PAPR before? (Yes or No)		If you have used a PAPR before, what specific activities do you use it for?	
Which hospital unit/department are you from? Please briefly describe your daily work activities.			
What is your experience with the following activities? (1: very experienced, 2: somewhat experienced; 3: inexperienced)			
Set up bedding/room			
Position patient			
Assess patient			
Hook up monitoring devices			
Insert IV			
Draw blood			
Start IV equipment			
Insert and set up foley			
Take urine sample			
Turn and bathe patient			
Suctioning			
Perform CPR			

Appendix B: User Evaluation Questionnaire Analysis of User Preference with N95 and Powered Air-Purifying Respirators in a healthcare work environment

Date: _____ Study ID: _____ 002 _____

BASIC INFORMATION

PLEASE PRINT AND COMPLETE ALL ENTRIES

NAME (LAST -- FIRST -- MIDDLE INITIAL)		JOB TITLE	
SEX <input type="checkbox"/> Male <input type="checkbox"/> Female	AGE (Years) Weight (lb.)	Have you used PAPRs?	YEARS OF WORK EXPERIENCE

List of Tested Respirators:

Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
Versflow	Koken	AIR MATE	MAX AIR	N95
Time Start:				
Time Stop:				

List of Exercises

Set up bedding/room

Position patient

Assess patient

Hook up monitoring devices

Insert IV

Draw blood

Start IV equipment

Insert and set up foley

Take urine sample

Turn and bathe patient

Suctioning

Perform CPR

Place a circle next to the following scales corresponding to your evaluation of each respirator.

General comfort				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Very poor	1 Very poor	1 Very poor	1 Very poor	1 Very poor
2 Poor	2 Poor	2 Poor	2 Poor	2 Poor
3 Acceptable	3 Acceptable	3 Acceptable	3 Acceptable	3 Acceptable
4 Good	4 Good	4 Good	4 Good	4 Good
5 Very good	5 Very good	5 Very good	5 Very good	5 Very good
6 Excellent	6 Excellent	6 Excellent	6 Excellent	6 Excellent

Inspiratory/expiratory effort				
Respirator	Respirator	Respirator	Respirator	Respirator
1	2	3	4	5

	Not noticeable	1		Not noticeable	1		Not noticeable	1		Not noticeable	1		Not noticeable	1
2														
3														
4														
5														
6														
7	Intolerable			Intolerable			Intolerable			Intolerable			Intolerable	

Overall breathing discomfort				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5

1	Not at all	1	Not at all	1	Not at all	1	Not at all	1	Not at all
2	Very slight	2	Very slight	2	Very slight	2	Very slight	2	Very slight
3	Slight —	3	Slight —	3	Slight —	3	Slight —	3	Slight —
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High —	5	High —	5	High —	5	High —	5	High —
e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable

Facial and body heat				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Not at all	1 Not at all	1 Not at all	1 Not at all	1 Not at all
2 Very slight	2 Very slight	2 Very slight	2 Very slight	2 Very slight
3 Slight —	3 Slight —	3 Slight —	3 Slight —	3 Slight —
4 Somewhat high	4 Somewhat high	4 Somewhat high	4 Somewhat high	4 Somewhat high
5 High —	5 High —	5 High —	5 High —	5 High —
6 Unbearable	6 Unbearable	6 Unbearable	6 Unbearable	6 Unbearable

Overall thermal comfort				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Coldest you've ever been	1 Coldest you've ever been	1 Coldest you've ever been	1 Coldest you've ever been	1 Coldest you've ever been

2	—	2	—	2	—	2	—	2	—
3	—	3	—	3	—	3	—	3	—
4	—	4	—	4	—	4	—	4	—
hot	Neither— 5 Nor cold	hot	Neither— 5 Nor cold	hot	Neither— 5 Nor cold	hot	Neither— 5 Nor cold	hot	Neither— 5 Nor cold
6	—	6	—	6	—	6	—	6	—
7	—	7	—	7	—	7	—	7	—
8	—	8	—	8	—	8	—	8	—
9	—	9	—	9	—	9	—	9	—
you	Hottest— 10 have ever been	you	Hottest— 10 have ever been	you	Hottest— 10 have ever been	you	Hottest— 10 have ever been	you	Hottest— 10 have ever been

Pressure or pain				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Not at all	1 Not at all	1 Not at all	1 Not at all	1 Not at all
2 Very slight	2 Very slight	2 Very slight	2 Very slight	2 Very slight
3 Slight —	3 Slight —	3 Slight —	3 Slight —	3 Slight —
4 Somewhat high	4 Somewhat high	4 Somewhat high	4 Somewhat high	4 Somewhat high
5 High —	5 High —	5 High —	5 High —	5 High —
6 Unbearable	6 Unbearable	6 Unbearable	6 Unbearable	6 Unbearable

Eye discomfort				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Not at all	1 Not at all	1 Not at all	1 Not at all	1 Not at all

2	Very slight	2	Very slight	2	Very slight	2	Very slight	2	Very slight
3	Slight –	3	Slight –	3	Slight –	3	Slight –	3	Slight –
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High –	5	High –	5	High –	5	High –	5	High –
e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl

Interference with wearing glasses/goggles/contact lenses (Don't complete the section if you don't wear one of those)				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1	1	1	1	1
2	2	2	2	2

3	Slight —	3	Slight —	3	Slight —	3	Slight —	3	Slight —
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High —	5	High —	5	High —	5	High —	5	High —
e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl

Clear line of vision				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Very poor	1 Very poor	1 Very poor	1 Very poor	1 Very poor
2 Poor	2 Poor	2 Poor	2 Poor	2 Poor
3 Acceptable	3 Acceptable	3 Acceptable	3 Acceptable	3 Acceptable
4 Good	4 Good	4 Good	4 Good	4 Good
5 Very good	5 Very good	5 Very good	5 Very good	5 Very good
6 Excellent	6 Excellent	6 Excellent	6 Excellent	6 Excellent

Difficulty of putting on				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5

1	Not at all	1	Not at all	1	Not at all	1	Not at all	1	Not at all
2	Very slight	2	Very slight	2	Very slight	2	Very slight	2	Very slight
3	Slight —	3	Slight —	3	Slight —	3	Slight —	3	Slight —
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High —	5	High —	5	High —	5	High —	5	High —
e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable

Difficulty to operate				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
1 Not at all	1 Not at all	1 Not at all	1 Not at all	1 Not at all
2 Very slight	2 Very slight	2 Very slight	2 Very slight	2 Very slight
3 Slight —	3 Slight —	3 Slight —	3 Slight —	3 Slight —
high 4 Somewhat	high 4 Somewhat	high 4 Somewhat	high 4 Somewhat	high 4 Somewhat
5 High —	5 High —	5 High —	5 High —	5 High —
e 6 Unbearabl	e 6 Unbearabl	e 6 Unbearabl	e 6 Unbearabl	e 6 Unbearabl

Mechanical interference with duties				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5

1	Not at all	1	Not at all	1	Not at all	1	Not at all	1	Not at all
2	Very slight	2	Very slight	2	Very slight	2	Very slight	2	Very slight
3	Slight —	3	Slight —	3	Slight —	3	Slight —	3	Slight —
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High —	5	High —	5	High —	5	High —	5	High —
e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable	e 6	Unbearable

Exertion				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
No exertion at all 6 7 8 9 10 11 12 13 14	No exertion at all 6 7 8 9 10 11 12 13 14	No exertion at all 6 7 8 9 10 11 12 13 14	No exertion at all 6 7 8 9 10 11 12 13 14	No exertion at all 6 7 8 9 10 11 12 13 14

15	—	15	—	15	—	15	—	15	—
16	—	16	—	16	—	16	—	16	—
17	—	17	—	17	—	17	—	17	—
18	—	18	—	18	—	18	—	18	—
19	—	19	—	19	—	19	—	19	—
20	Maximal— exertion	20	Maximal— exertion	20	Maximal— exertion	20	Maximal— exertion	20	Maximal— exertion

How many hours do you think you could wear this respirator continuously?				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5

Perceived efficiency against biological hazards				
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5

1	Not at all	1	Not at all	1	Not at all	1	Not at all	1	Not at all
2	Very slight	2	Very slight	2	Very slight	2	Very slight	2	Very slight
3	Slight —	3	Slight —	3	Slight —	3	Slight —	3	Slight —
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High —	5	High —	5	High —	5	High —	5	High —
6	Complete	6	Complete	6	Complete	6	Complete	6	Complete

Study ID: _____

Overall assessment					
Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5	
1	Not at all	1	Not at all	1	Not at all
2	Very slight	2	Very slight	2	Very slight

3	Slight –	3	Slight –	3	Slight –	3	Slight –	3	Slight –
high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat	high 4	Somewhat
5	High –	5	High –	5	High –	5	High –	5	High –
e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl	e 6	Unbearabl

BEFORE TEST

RESPIRATOR TYPE	VITAL SIGNS			EXPERIENCE OF USING ANY TYPE OF RESPIRATOR BEFORE
	Heart Rate (/mins)	Respiratory Rate(/mins)	SpO2 (%)	
Respirator 1				
Respirator 2				
Respirator 3				
Respirator 4				
Respirator 5				

AFTER TEST

RESPIRATOR TYPE	VITAL SIGNS			SPEECH INTELLIGIBILITY TEST (MRT)
	Heart Rate (/mins)	Respiratory Rate(/mins)	SpO2 (%)	
Respirator 1				

Respirator 2				
Respirator 3				
Respirator 4				
Respirator 5				

QUESTIONNAIRE SURVEY

Evaluation on preference and acceptability comparing low flow rate with high flow rate for low-risk tasks					
Questions	Versflow	Koken	AIR MATE	MAX AIR	N95
When performing the following tasks (deemed low risks), do you prefer a lower flow rate for the following PAPR? (yes and no for each PAPR model)					<i>N/A</i>
When performing the following tasks (deemed low risks), do you think a lower flow rate is more acceptable? (yes and no for each PAPR model)					<i>N/A</i>
When performing the following tasks (deemed low risks), do you think a lower flow rate is more effective? (yes and no for each PAPR model)					<i>N/A</i>

Evaluation on preference and acceptability comparing N95 FFR with evaluated PAPRs for low-risk tasks					
Questions	Versflow	Koken	AIR MATE	MAX AIR	N95
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you prefer to use?					
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more acceptable?					
When performing the following tasks (deemed low risks), which type of the following 5 respirators do you think is more effective?					

Evaluation on preference and acceptability comparing N95 FFR with evaluated PAPRs for high-risk tasks					
Questions	Versflow	Koken	AIR MATE	MAX AIR	N95
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you prefer to use?					
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more acceptable?					
When performing the following tasks (deemed high risks), which type of the following 5 respirators do you think is more effective?					

Other symptoms

<i>Somatic complaints for respirators (Yes and No for each symptom)</i>	Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
<i>Nausea</i>					
<i>Dizziness or difficult concentrating</i>					
<i>Unusual smell/odor</i>					
<i>Eye irritation signs (if any)</i>					

List of Tested Respirators:

Respirator 1	Respirator 2	Respirator 3	Respirator 4	Respirator 5
Versflow	Koken	AIR MATE	MAX AIR	N95

Other Comments:

Appendix C: Raw Data

C.1. Evaluation Categories

N95	Subject 001	Subject 002	Subject 003	Subject 004	Subject 005	Subject 006	Subject 007	Subject 008
General comfort	2	3	2	5	2	3	2	3
Inspiratory/expiratory effort	4	5	4	2	6	2	2	5
Overall breathing discomfort	2	4	4	2	5	2	2	4
Facial and body heat	2	4	4	2	5	2	4	5
Overall thermal comfort	6	7	7	6	7	5	5	10
Pressure or pain	4	3	3	2	4	2	3	1
Eye discomfort	1	1	1	1	1	1	1	1
Interference with wearing glasses/goggles/contacts	1	2	N/A	1	1	1	2	N/A
Clear line of vision	5	5	6	6	5	5	5	5
Difficulty of putting on	1	2	1	1	1	2	1	1
Difficulty to operate	1	1	1	1	1	1	1	1
Mechanical interference with duties	2	1	1	1	1	1	1	1
Exertion	14	14	9	6	12	7	13	6
How many hours could you wear this respiratory continuously?	1	0.5	1	12	2	2	1	1
Perceived efficiency against biological hazards	3	4	4	6	6	3	4	4
Overall assessment	3	5	4	2	4	2	4	1

Subject 009	Subject 010	Subject 011	Subject 012	Subject 013	Subject 014	Subject 015	Subject 016
3	3	5	1	3	2	3	3
3	1	1	7	5	5	3	2
2	2	1	5	3	5	3	2
4	3	2	6	5	5	4	2
7	5	5	8	7	8	6	6
3	1	2	4	1	1	1	3
3	1	1	1	1	4	1	1
3	1	1	1	N/A	5	N/A	3
4	4	5	6	5	2	6	4
3	1	1	1	2	1	1	1
1	1	1	1	1	3	1	1
2	1	1	1	1	4	1	2
11	6	6	10	10	10	7	8
2	1	2	0.05	1.5	1	1	5
2	3	3	4	2	1	3	4
3	4	2	6	3	4	4	4

Versflow (PAPR)	Subject 001	Subject 002	Subject 003	Subject 004	Subject 005	Subject 006	Subject 007	Subject 008
General comfort	2	1	2	3	1	3	4	5
Inspiratory/expiratory effort	4	5	4	2	2	2	2	1
Overall breathing discomfort	3	4	4	2	5	1	2	1
Facial and body heat	1	5	5	2	5	2	3	1
Overall thermal comfort	3	7	8	5	8	5	6	2
Pressure or pain	5	1	2	1	3	3	2	1
Eye discomfort	1	1	3	1	3	2	2	1
Interference with wearing glasses/goggles/contacts	1	4	N/A	1	2	2	3	N/A
Clear line of vision	3	3	2	4	4	3	4	4
Difficulty of putting on	5	3	4	4	3	3	2	1
Difficulty to operate	5	4	3	4	5	3	2	1
Mechanical interference with duties	5	4	4	3	5	3	2	2
Exertion	17	12	10	8	17	8	9	6
How many hours could you wear this respiratory continuously?	1	1	2	2	0.25	2	3	2
Perceived efficiency against biological hazards	3	3	5	6	3	3	3	3
Overall assessment	5	5	5	3	5	3	2	1

Subject 009	Subject 010	Subject 011	Subject 012	Subject 013	Subject 014	Subject 015	Subject 016
5	2	2	3	3	4	2	4
2	2	1	4	1	1	2	1
2	2	2	1	1	1	1	1
3	4	1	1	2	1	3	1
6	7	5	4	5	5	5	5
1	1	3	1	1	1	1	2
1	3	1	1	1	1	1	1
1	3	3	1	N/A	1	N/A	2
4	2	2	3	3	4	5	4
3	5	4	3	4	3	4	2
3	5	4	2	3	1	2	2
2	5	4	2	3	1	4	3
8	6	9	10	8	8	6	8
4	0.25	1	0.2	3	1.5	2	6
5	3	4	5	4	2	3	5
2	5	5	2	2	2	3	5

MAX AIR (PAPR)	Subject 001	Subject 002	Subject 003	Subject 004	Subject 005	Subject 006	Subject 007	Subject 008
General comfort	5	3	5	5	6	4	5	5
Inspiratory/expiratory effort	1	1	2	2	1	1	2	1
Overall breathing discomfort	1	2	1	2	1	1	1	1
Facial and body heat	2	2	3	2	1	2	2	1
Overall thermal comfort	5	4	7	5	5	5	5	4
Pressure or pain	3	1	1	1	1	2	1	1
Eye discomfort	1	1	1	1	1	2	3	1
Interference with wearing glasses/goggles/contacts	1	2	N/A	1	1	1	2	N/A
Clear line of vision	4	3	6	5	6	4	4	5
Difficulty of putting on	4	3	3	2	1	2	2	3
Difficulty to operate	3	3	1	2	1	2	2	1
Mechanical interference with duties	2	3	2	1	1	2	2	4
Exertion	10	12	8	6	6	7	7	6
How many hours could you wear this respiratory continuously?	2	2	4	8	4	3	4	2
Perceived efficiency against biological hazards	4	4	6	5	5	3	3	3
Overall assessment	3	3	2	2	1	2	2	1

Subject 009	Subject 010	Subject 011	Subject 012	Subject 013	Subject 014	Subject 015	Subject 016
4	3	3	2	3	4	4	4
2	2	1	2	1	2	1	1
1	2	1	1	1	1	1	1
2	2	3	1	2	4	2	1
5	5	6	5	5	7	5	5
2	1	1	5	2	1	1	2
1	1	1	1	1	1	1	1
1	1	1	2	N/A	2	N/A	2
4	4	5	2	3	4	5	5
2	3	1	2	4	1	2	1
2	3	1	1	3	1	2	1
3	2	2	4	4	1	3	2
8	6	7	13	10	6	8	8
4	1	1	0.15	2	1.5	2	6
5	3	5	2	4	2	3	4
2	5	2	2	3	4	3	4

AIR MATE (PAPR)	Subject 001	Subject 002	Subject 003	Subject 004	Subject 005	Subject 006	Subject 007	Subject 008
General comfort	3	4	4	3	4	3	5	5
Inspiratory/expiratory effort	1	4	2	2	3	1	2	1
Overall breathing discomfort	1	3	2	2	2	1	2	1
Facial and body heat	1	3	3	4	2	2	2	1
Overall thermal comfort	5	5	5	7	5	5	5	4
Pressure or pain	3	1	2	1	1	2	1	1
Eye discomfort	1	1	2	1	1	2	1	1
Interference with wearing glasses/goggles/contacts	1	2	N/A	1	1	2	2	N/A
Clear line of vision	3	3	3	5	4	3	3	2
Difficulty of putting on	3	3	3	4	5	2	2	3
Difficulty to operate	3	3	2	4	2	2	1	1
Mechanical interference with duties	3	2	2	3	2	3	1	1
Exertion	12	12	9	10	8	7	8	6
How many hours could you wear this respirator continuously?	1	2	3	1	2	3	4	1
Perceived efficiency against biological hazards	3	4	5	5	5	3	3	3
Overall assessment	2	3	3	3	3	3	2	1

Subject 009	Subject 010	Subject 011	Subject 012	Subject 013	Subject 014	Subject 015	Subject 016
3	4	3	2	3	4	3	4
3	1	1	4	1	3	3	1
3	1	1	1	1	1	3	1
3	1	1	2	2	3	4	1
6	5	5	3	5	6	6	5
2	1	1	1	1	1	1	2
4	1	1	4	1	1	1	1
4	1	1	4	N/A	3	N/A	2
3	5	6	4	3	3	5	4
4	2	2	2	3	1	2	2
3	2	2	1	3	1	2	2
4	1	4	4	4	1	3	3
13	6	6	10	8	6	7	8
1	2	1	0.2	3	1.5	2	4
5	3	4	1	4	1	3	4
3	5	2	4	2	3	4	4